

Effects of incorporation of hydroxyapatite and fluoroapatite nanobioceramics into conventional glass ionomer cements (GIC)

Moshaverinia, A.^a, Ansari, S.^a, Moshaverinia, M.^b, Roohpour, N.^a, Darr, J.A.^c, Rehman, I.^a

^a Department of Materials, Interdisciplinary Research Centre in Biomedical Materials, Queen Mary University of London, Mile End Road, London, E1 4NS, United Kingdom

^b Department of Oral Medicine, Mashhad Dental School, Mashhad University of Medical Sciences, P.O. Box 91730-9184, Mashhad, Iran

^c Department of Chemistry, University College London, Christopher Ingold Laboratories, 2 Gordon Street, London, WC1H 0AJ, United Kingdom

[View references \(4\)](#)

Abstract

Hydroxyapatite (HA) has excellent biological behavior, and its composition and crystal structure are similar to the apatite in the human dental structure and skeletal system; a number of researchers have attempted to evaluate the effect of the addition of HA powders to restorative dental materials. In this study, nanohydroxy and fluoroapatite were synthesized using an ethanol based sol-gel technique. The synthesized nanoceramic particles were incorporated into commercial glass ionomer powder (Fuji II GC) and were characterized using Fourier transform infrared and Raman spectroscopy, X-ray diffraction and scanning electron microscopy. Compressive, diametral tensile and biaxial flexural strengths of the modified glass ionomer cements were evaluated. The effect of nanohydroxyapatite and fluoroapatite on the bond strength of glass ionomer cement to dentin was also investigated. Results showed that after 1 and 7 days of setting, the nanohydroxyapatite/fluoroapatite added cements exhibited higher compressive strength (177-179 MPa), higher diametral tensile strength (16-20 MPa) and higher biaxial flexural strength (26-28 MPa) as compared with the control group (160 MPa in CS, 14 MPa in DTS and 18 MPa in biaxial flexural strength). The experimental cements also exhibited higher bond strength to dentin after 7 and 20 days of storage in distilled water. It was concluded that glass ionomer cements containing nanobioceramics are promising restorative dental materials with both improved mechanical properties and improved bond strength to dentin. © 2007 Acta Materialia Inc.

Reaxys Database Information

Author keywords

Glass ionomer cement; Mechanical properties; Nanofluoroapatite; Nanohydroxyapatite; Sol-gel synthesis

Indexed Keywords

EMTREE drug terms: alcohol; apatite; fluorapatite; glass ionomer; hydroxyapatite; water

EMTREE medical terms: article; ceramics; chemical composition; compressive strength; controlled study; crystal structure; dentin; infrared spectroscopy; powder; priority journal; Raman spectrometry; scanning electron microscopy; skeleton; synthesis; tensile strength; tooth; X ray diffraction

MeSH: Apatites; Biomechanics; Durapatite; Glass Ionomer Cements; Materials Testing; Microscopy, Electron, Scanning; Nanostructures; Nanotechnology; Spectroscopy, Fourier Transform Infrared; Surface Properties; X-Ray Diffraction

Medline is the source for the MeSH terms of this document.

Chemicals and CAS Registry Numbers: alcohol, 64-17-0; apatite, 74476-28-7; fluorapatite, 12020-40-0, 1306-00-4, 08051-82-4; hydroxyapatite, 1306-06-0, 01198-94-8; water, 7732-18-0; Apatites; Durapatite, 1306-06-0; fluorapatite, 1306-00-4; Glass Ionomer Cements